
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000922
13:01 10/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 1:30
STOP BUTTON PRESSED
NOT STERILE
END TIME 1:30

Digital Signature #
3DDFD8AEE0DBC5F7
