

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3166

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
2:00 174.6 F 2.2  
4:00 193.0 F 6.7  
6:00 211.7 F 11.3  
8:00 225.9 F 6.3  
10:00 232.5 F 11.6  
12:00 235.6 F 9.4  
14:00 238.3 F 9.7  
16:00 248.9 F 15.1  
18:00 260.8 F 21.8  
20:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.5 F 29.5  
2:00 272.5 F 29.9  
3:00 271.2 F 29.3  
4:00 272.7 F 29.7  
  
Min 271.2 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 169.5 F 0.0  
5:00 228.7 F 0.0  
10:00 196.5 F 0.0  
15:00 194.8 F 0.0  
20:00 191.2 F 0.0  
25:00 188.8 F 0.0  
30:00 186.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 21:21  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:00:33  
CYCLE COMPLETE

11/29/2023 12:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_