
Midmark M11 - v1.0.5
Total Cycles: 4034

Name: RITTER2
Sterilizer ID: V2313228
10/16/2024 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 151.1 F 1.5
4:00 153.3 F 4.5
6:00 178.4 F 8.4
8:00 201.4 F 12.9
10:00 212.7 F 6.6
12:00 225.8 F 11.4
14:00 229.8 F 8.1
16:00 240.2 F 12.3
18:00 243.0 F 12.7
20:00 255.4 F 19.1
22:00 266.4 F 25.8
24:00 272.4 F 30.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.9
1:00 273.2 F 30.6
2:00 272.1 F 30.0
3:00 271.7 F 30.2
4:00 273.0 F 30.5

Min 271.4 F 29.8
Max 273.3 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 209.9 F 0.0
10:00 212.0 F 0.0
15:00 204.5 F 0.0
20:00 193.6 F 0.0
25:00 185.8 F 0.0
30:00 181.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:30
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

10/16/2024 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____