

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1579

Name: RITTER2  
Sterilizer ID: V2313228  
10/11/2022 09:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
2:00 181.5 F 0.1  
4:00 194.5 F 1.0  
6:00 219.0 F 7.6  
8:00 232.1 F 10.0  
10:00 241.5 F 12.7  
12:00 242.4 F 11.9  
14:00 242.7 F 11.7  
16:00 258.1 F 19.8  
18:00 271.0 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.4 F 28.8  
2:00 272.7 F 29.3  
3:00 271.7 F 28.3  
4:00 272.9 F 29.2  
  
Min 271.5 F 28.3  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 222.6 F 0.0  
5:00 240.4 F 0.0  
Dry Cycle Aborted

FILLING: 1:30  
HEATING: 18:39  
STERILIZING: 4:00  
VENTING: 2:44  
DRYING: 7:29  
TOTAL CYCLE: 00:34:22

10/11/2022 10:22 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_