
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 000302
15:11 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:35
PRESSURIZING 3:40
STERILIZING 4:50
136.0 C 229kPa 4:50
136.4 C 233kPa 5:20
136.2 C 232kPa 5:50
136.3 C 232kPa 6:20
136.3 C 232kPa 6:50
136.4 C 231kPa 7:20
136.4 C 233kPa 7:50
136.2 C 232kPa 8:20
136.2 C 232kPa 8:20
VENTING 8:20
AIR DRYING 9:13
CYCLE COMPLETE 9:14

Digital Signature #
45AE007B3050BDFD
