
Midmark M11 - v1.0.0
Total Cycles: 3783

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2023 03:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 155.9 F 1.5
4:00 177.5 F 5.1
6:00 199.1 F 9.4
8:00 220.5 F 9.3
10:00 224.9 F 8.4
12:00 236.6 F 13.2
14:00 239.3 F 10.3
16:00 236.0 F 8.2
18:00 249.7 F 14.8
20:00 260.9 F 21.1
22:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.3 F 28.5
2:00 272.1 F 28.8
3:00 272.4 F 28.8
4:00 271.7 F 28.7

Min 270.7 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.7 F 0.0
5:00 205.8 F 0.0
10:00 246.9 F 0.0
15:00 228.2 F 0.0
20:00 212.0 F 0.0
25:00 190.3 F 0.0
30:00 183.6 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

09/15/2023 04:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____