

-----  
Midmark M11 - v1.0.5  
Total Cycles: 547

Name: RITTER2  
Sterilizer ID: V2313228  
11/15/2021 07:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 89.8 F 0.0  
4:00 124.0 F 0.0  
6:00 164.2 F 0.1  
8:00 181.6 F 0.2  
10:00 200.4 F 0.4  
12:00 214.2 F 2.7  
14:00 222.7 F 6.1  
16:00 230.3 F 9.6  
18:00 238.0 F 12.1  
20:00 239.7 F 11.1  
22:00 236.7 F 9.2  
24:00 238.6 F 9.6  
26:00 249.2 F 15.1  
28:00 261.0 F 21.7  
30:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.3  
1:00 272.4 F 29.7  
2:00 272.7 F 29.6  
3:00 272.4 F 29.5  
4:00 272.7 F 29.3

Min 271.8 F 29.1  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 237.0 F 0.0  
10:00 220.1 F 0.0  
15:00 204.1 F 0.0  
20:00 217.5 F 0.0  
25:00 230.2 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 31:08  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 01:09:01  
CYCLE COMPLETE

11/15/2021 09:24 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02