

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000686  
8:58 15/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:37  
STERILIZING 3:36  
136.4 C 231kPa 3:36  
136.4 C 233kPa 4:07  
136.4 C 232kPa 4:37  
136.4 C 233kPa 5:07  
136.4 C 233kPa 5:37  
136.4 C 233kPa 6:07  
136.3 C 229kPa 6:37  
136.4 C 230kPa 7:07  
136.4 C 230kPa 7:07  
VENTING 7:07  
AIR DRYING 8:00  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
B6F608C5FDEB368E  
-----