

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3002

Name: RITTER2  
Sterilizer ID: V2313228  
10/12/2023 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.4 F 0.0  
2:00 181.2 F 3.9  
4:00 201.3 F 9.5  
6:00 219.4 F 6.7  
8:00 232.9 F 11.3  
10:00 235.5 F 9.3  
12:00 237.2 F 9.2  
14:00 250.6 F 16.0  
16:00 263.5 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.7 F 29.5  
2:00 271.4 F 28.8  
3:00 272.5 F 29.7  
4:00 272.0 F 29.1  
  
Min 271.3 F 28.7  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.7 F 0.0  
5:00 228.1 F 0.0  
10:00 213.4 F 0.0  
15:00 203.9 F 0.0  
20:00 201.2 F 0.0  
25:00 206.2 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 17:58  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:12  
CYCLE COMPLETE

10/12/2023 12:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_