
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001198
12:38 28/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:54
STERILIZING 5:09
136.0 C 221kPa 5:09
136.3 C 225kPa 5:39
136.4 C 226kPa 6:09
136.4 C 225kPa 6:39
136.1 C 223kPa 7:09
136.1 C 223kPa 7:39
136.4 C 226kPa 8:09
136.4 C 225kPa 8:39
136.4 C 225kPa 8:39
VENTING 8:39
AIR DRYING 9:39
CYCLE COMPLETE 15:41

Digital Signature #
12017601B2CC6056
