
Midmark M11 - v1.0.5
Total Cycles: 707

Name: RITTER2
Sterilizer ID: V2313228
01/19/2022 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 178.6 F 3.8
4:00 199.0 F 9.1
6:00 218.1 F 7.6
8:00 229.2 F 10.8
10:00 232.7 F 8.6
12:00 238.7 F 10.3
14:00 248.0 F 14.9
16:00 260.6 F 22.0
18:00 272.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.3 F 29.4
2:00 271.9 F 29.7
3:00 272.0 F 29.3
4:00 272.0 F 29.8

Min 271.4 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 217.8 F 0.0
10:00 202.8 F 0.0
15:00 203.2 F 0.0
20:00 202.3 F 0.0
25:00 208.8 F 0.0
30:00 198.6 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:29
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:35
CYCLE COMPLETE

01/19/2022 03:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____