
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002245
12:03 12/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:44
STERILIZING 3:44
136.4 C 222kPa 3:44
136.2 C 224kPa 4:14
136.4 C 225kPa 4:44
136.3 C 223kPa 5:14
136.3 C 224kPa 5:44
136.4 C 223kPa 6:14
136.2 C 222kPa 6:44
136.4 C 225kPa 7:14
136.4 C 225kPa 7:14
VENTING 7:14
AIR DRYING 8:07
CYCLE COMPLETE 38:07

Digital Signature #
DFE767027A0134A8
