

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1021

Name: RITTER 1  
Sterilizer ID: V2178198  
07/15/2021 01:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
2:00 188.6 F 3.8  
4:00 205.7 F 8.7  
6:00 227.0 F 8.5  
8:00 236.1 F 9.9  
10:00 234.6 F 8.2  
12:00 244.1 F 12.1  
14:00 246.1 F 13.1  
16:00 258.6 F 19.9  
18:00 269.7 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.6  
2:00 271.5 F 28.4  
3:00 272.4 F 28.6  
4:00 271.2 F 28.1  
  
Min 271.2 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 229.4 F 0.0  
10:00 221.8 F 0.0  
15:00 230.5 F 0.0  
20:00 223.7 F 0.0  
25:00 225.1 F 0.0  
30:00 220.8 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 18:55  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

07/15/2021 02:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_