

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2793

Name: RITTER2  
Sterilizer ID: V2313228  
08/21/2023 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 67.1 F 0.0  
2:00 78.8 F 0.0  
4:00 79.6 F 0.0  
6:00 104.8 F 0.0  
8:00 149.7 F 0.4  
10:00 180.9 F 2.9  
12:00 194.9 F 6.8  
14:00 211.5 F 10.6  
16:00 221.2 F 6.2  
18:00 230.3 F 9.2  
20:00 238.9 F 11.0  
22:00 241.0 F 11.5  
24:00 238.8 F 9.9  
26:00 252.9 F 16.9  
28:00 264.7 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.1 F 28.6  
2:00 272.7 F 29.0  
3:00 272.7 F 29.4  
4:00 271.7 F 28.6

Min 271.6 F 28.3  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 234.3 F 0.0  
10:00 187.7 F 0.0  
15:00 187.7 F 0.0  
20:00 186.8 F 0.0  
25:00 205.8 F 0.0  
30:00 217.1 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:50  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 01:07:36  
CYCLE COMPLETE

08/21/2023 11:58 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_