

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1454

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2021 10:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 173.7 F 2.4  
4:00 187.9 F 5.9  
6:00 199.5 F 10.3  
8:00 216.6 F 4.8  
10:00 227.4 F 8.9  
12:00 237.2 F 11.5  
14:00 239.2 F 10.8  
16:00 237.1 F 8.6  
18:00 250.0 F 15.2  
20:00 261.6 F 21.8  
22:00 271.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.1  
2:00 272.0 F 28.4  
3:00 271.8 F 28.7  
4:00 272.6 F 29.0

Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 223.2 F 0.0  
10:00 240.6 F 0.0  
15:00 228.9 F 0.0  
20:00 222.7 F 0.0  
25:00 222.1 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:35  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

12/14/2021 11:48 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_