
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000117
14:56 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:26
STERILIZING 5:35
136.0 C 222kPa 5:35
136.3 C 224kPa 6:05
136.3 C 225kPa 6:35
136.3 C 226kPa 7:05
136.4 C 227kPa 7:35
136.2 C 225kPa 8:05
136.3 C 223kPa 8:35
136.4 C 226kPa 9:05
136.4 C 226kPa 9:05
VENTING 9:05
AIR DRYING 10:10
CYCLE COMPLETE 25:10

Digital Signature #
933A0CB2796D9859
