
Midmark M11 - v1.0.0
Total Cycles: 4381

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2024 10:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.2 F 0.0
2:00 139.8 F 1.4
4:00 156.6 F 4.1
6:00 180.3 F 7.6
8:00 201.0 F 11.5
10:00 216.5 F 3.6
12:00 222.8 F 7.5
14:00 233.0 F 11.7
16:00 232.9 F 7.7
18:00 245.0 F 13.1
20:00 242.0 F 11.3
22:00 253.7 F 17.3
24:00 264.3 F 23.5
26:00 271.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.5 F 28.7
2:00 271.2 F 28.6
3:00 272.4 F 28.6
4:00 271.6 F 28.7

Min 271.0 F 28.2
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 226.3 F 0.0
10:00 233.2 F 0.0
15:00 218.4 F 0.0
20:00 223.7 F 0.0
25:00 221.3 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:48
CYCLE COMPLETE

03/15/2024 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____