
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000201
11:20 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:45
STERILIZING 5:20
136.0 C 228kPa 5:20
135.9 C 229kPa 5:50
136.1 C 231kPa 6:20
136.4 C 233kPa 6:50
136.3 C 231kPa 7:20
136.3 C 231kPa 7:50
136.4 C 233kPa 8:20
136.4 C 233kPa 8:50
136.4 C 233kPa 8:50
VENTING 8:50
AIR DRYING 9:53
CYCLE COMPLETE 24:53

Digital Signature #
CDE3D2475857732B
