
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000956
9:05 10/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:45
STERILIZING 3:43
136.0 C 220kPa 3:43
136.2 C 222kPa 4:13
136.4 C 222kPa 4:43
136.4 C 223kPa 5:13
136.4 C 222kPa 5:43
136.4 C 225kPa 6:13
136.2 C 223kPa 6:43
136.0 C 221kPa 7:13
136.0 C 221kPa 7:13
VENTING 7:13
AIR DRYING 8:03
CYCLE COMPLETE 8:20

Digital Signature #
231540FF19DF5704
