
Midmark M11 - v1.0.5
Total Cycles: 1259

Name: RITTER2
Sterilizer ID: V2313228
07/06/2022 07:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.4 F 0.0
2:00 73.7 F 0.0
4:00 92.5 F 0.0
6:00 138.0 F 0.0
8:00 153.7 F 0.2
10:00 185.3 F 1.7
12:00 197.8 F 5.4
14:00 213.1 F 9.0
16:00 225.3 F 12.8
18:00 230.1 F 8.2
20:00 241.3 F 13.0
22:00 239.7 F 10.8
24:00 236.9 F 9.2
26:00 250.6 F 15.9
28:00 262.0 F 22.5
30:00 272.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.8
3:00 272.7 F 29.5
4:00 271.6 F 28.7

Min 271.4 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 201.6 F 0.0
10:00 226.1 F 0.0
15:00 227.9 F 0.0
20:00 231.4 F 0.0
25:00 230.8 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:30
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

07/06/2022 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____