
Midmark M11 - v1.0.5
Total Cycles: 175

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.0 F 0.0
2:00 145.4 F 2.3
4:00 164.2 F 6.1
6:00 189.9 F 10.6
8:00 209.4 F 4.8
10:00 222.9 F 9.4
12:00 234.1 F 10.1
14:00 240.1 F 12.1
16:00 239.7 F 10.8
18:00 252.9 F 17.1
20:00 263.4 F 23.7
22:00 272.1 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 271.5 F 29.6
2:00 272.2 F 29.6
3:00 271.2 F 29.0
4:00 271.5 F 29.7

Min 271.0 F 29.0
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 219.6 F 0.0
10:00 218.8 F 0.0
15:00 218.2 F 0.0
20:00 215.6 F 0.0
25:00 214.7 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

07/07/2021 02:17 PM
ACCEPTED: X
REJECTED:
Operator: U01