

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2216

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2023 01:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.8 F 0.0  
2:00 180.0 F 4.0  
4:00 195.5 F 8.7  
6:00 216.2 F 8.3  
8:00 231.0 F 10.2  
10:00 232.6 F 7.9  
12:00 240.4 F 10.8  
14:00 247.1 F 14.3  
16:00 260.3 F 21.6  
18:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.7  
2:00 271.6 F 29.0  
3:00 272.3 F 30.0  
4:00 272.2 F 29.5  
  
Min 271.4 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 237.8 F 0.0  
10:00 221.2 F 0.0  
15:00 220.4 F 0.0  
20:00 221.3 F 0.0  
25:00 221.5 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 18:34  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:40  
CYCLE COMPLETE

04/04/2023 02:19 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_