

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4376

Name: RITTER 1  
Sterilizer ID: V2178198  
03/12/2024 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.1 F 0.0  
2:00 157.8 F 1.9  
4:00 165.4 F 5.0  
6:00 182.5 F 8.3  
8:00 203.0 F 12.7  
10:00 224.0 F 5.9  
12:00 230.7 F 10.4  
14:00 233.5 F 7.8  
16:00 242.6 F 12.2  
18:00 239.9 F 10.3  
20:00 252.2 F 16.4  
22:00 262.9 F 22.6  
24:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 273.0 F 29.2  
2:00 271.1 F 28.7  
3:00 272.7 F 29.1  
4:00 271.4 F 28.9

Min 271.0 F 28.6  
Max 273.1 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 212.5 F 0.0  
10:00 189.1 F 0.0  
15:00 184.4 F 0.0  
20:00 182.7 F 0.0  
25:00 181.4 F 0.0  
30:00 180.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:58  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:03:48  
CYCLE COMPLETE

03/12/2024 11:09 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_