
Midmark M11 - v1.0.5
Total Cycles: 3243

Name: RITTER2
Sterilizer ID: V2313228
12/21/2023 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.5 F 0.0
4:00 82.6 F 0.4
6:00 115.9 F 1.7
8:00 150.8 F 4.6
10:00 172.6 F 7.8
12:00 194.3 F 12.0
14:00 203.3 F 4.4
16:00 221.8 F 8.2
18:00 233.2 F 12.1
20:00 233.2 F 8.0
22:00 243.6 F 13.0
24:00 241.5 F 11.1
26:00 254.0 F 17.5
28:00 265.1 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 28.9
2:00 272.1 F 29.3
3:00 271.2 F 28.5
4:00 272.4 F 29.2

Min 271.1 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 234.1 F 0.0
10:00 233.5 F 0.0
15:00 219.3 F 0.0
20:00 226.2 F 0.0
25:00 229.7 F 0.0
30:00 215.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

12/21/2023 10:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____