
Midmark M11 - v1.0.0
Total Cycles: 4238

Name: RITTER 1
Sterilizer ID: V2178198
01/26/2024 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 162.8 F 2.8
4:00 178.3 F 6.2
6:00 200.3 F 10.4
8:00 221.4 F 4.8
10:00 224.8 F 8.9
12:00 237.1 F 12.0
14:00 239.5 F 10.7
16:00 236.0 F 8.6
18:00 250.0 F 15.1
20:00 261.2 F 21.4
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.1 F 28.7
2:00 272.3 F 29.0
3:00 272.3 F 29.3
4:00 271.4 F 28.7

Min 271.4 F 28.6
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.6 F 0.0
5:00 241.5 F 0.0
10:00 201.9 F 0.0
15:00 225.3 F 0.0
20:00 192.8 F 0.0
25:00 203.6 F 0.0
30:00 207.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:17
CYCLE COMPLETE

01/26/2024 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____