
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002592
8:21 24/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 2:03
STERILIZING 3:14
136.4 C 223kPa 3:14
136.0 C 220kPa 3:44
136.1 C 221kPa 4:14
136.4 C 224kPa 4:44
136.5 C 223kPa 5:14
136.0 C 221kPa 5:44
136.4 C 224kPa 6:14
136.4 C 221kPa 6:44
136.4 C 221kPa 6:44
VENTING 6:44
AIR DRYING 7:37
CYCLE COMPLETE 23:31

Digital Signature #
DA5E5004F33C0F42
