
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007905
11:13 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:06
STERILIZING 3:12
136.0 C 220kPa 3:12
136.4 C 224kPa 3:42
136.1 C 223kPa 4:12
136.4 C 224kPa 4:42
136.3 C 221kPa 5:12
135.9 C 220kPa 5:42
136.2 C 222kPa 6:12
136.4 C 224kPa 6:42
136.4 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:34
CYCLE COMPLETE 31:43

Digital Signature #
9E06025D0A741AA0
