
Midmark M11 - v1.0.0
Total Cycles: 3118

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2023 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 167.8 F 0.2
4:00 182.5 F 3.8
6:00 200.1 F 8.2
8:00 221.0 F 13.1
10:00 225.6 F 7.0
12:00 236.1 F 11.6
14:00 235.0 F 8.5
16:00 241.6 F 11.1
18:00 245.9 F 13.2
20:00 257.9 F 19.5
22:00 268.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 28.9
2:00 271.8 F 28.2
3:00 272.0 F 28.7
4:00 272.4 F 28.8

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 210.2 F 0.0
10:00 215.1 F 0.0
15:00 218.8 F 0.0
20:00 226.2 F 0.0
25:00 230.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:13
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

04/14/2023 05:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____