
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007146
6:51 05/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:12
136.4 C 221kPa 4:12
136.4 C 220kPa 4:42
136.0 C 221kPa 5:12
136.2 C 222kPa 5:42
135.7 C 219kPa 6:12
136.4 C 222kPa 6:42
136.1 C 217kPa 7:12
136.4 C 223kPa 7:42
136.4 C 223kPa 7:42
VENTING 7:42
AIR DRYING 8:29
CYCLE COMPLETE 8:45

Digital Signature #
B8A53A6E3181EA44
