

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000674  
11:59 21/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:40  
PRESSURIZING 3:41  
STERILIZING 4:51  
136.5 C 220kPa 4:51  
136.3 C 224kPa 5:21  
136.4 C 225kPa 5:51  
136.2 C 223kPa 6:21  
136.4 C 225kPa 6:51  
136.4 C 224kPa 7:21  
136.2 C 223kPa 7:51  
136.3 C 224kPa 8:21  
136.3 C 224kPa 8:21  
VENTING 8:21  
AIR DRYING 9:17  
CYCLE COMPLETE 9:28

-----  
Digital Signature #  
AAF37B41D8DBD522  
-----