
Midmark M11 - v1.0.0
Total Cycles: 3289

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2023 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 169.3 F 0.0
4:00 172.2 F 0.8
6:00 194.0 F 5.1
8:00 211.3 F 9.5
10:00 221.7 F 8.1
12:00 229.9 F 8.5
14:00 241.2 F 13.3
16:00 240.3 F 10.7
18:00 237.1 F 9.0
20:00 251.3 F 15.6
22:00 262.8 F 22.2
24:00 272.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.7
2:00 271.8 F 28.0
3:00 272.0 F 28.5
4:00 272.5 F 28.6

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 213.0 F 0.0
10:00 215.0 F 0.0
15:00 216.8 F 0.0
20:00 216.7 F 0.0
25:00 214.6 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

05/26/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____