
Midmark M11 - v1.0.5
Total Cycles: 214

Name: RITTER2
Sterilizer ID: V2313228
07/15/2021 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.2 F 0.0
2:00 198.3 F 4.5
4:00 222.3 F 11.3
6:00 231.3 F 8.9
8:00 237.4 F 9.6
10:00 241.1 F 10.7
12:00 250.2 F 15.6
14:00 263.3 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.7 F 29.3
2:00 272.5 F 29.6
3:00 271.9 F 28.7
4:00 272.7 F 29.4

Min 271.7 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 216.0 F 0.0
10:00 208.8 F 0.0
15:00 209.0 F 0.0
20:00 204.5 F 0.0
25:00 201.4 F 0.0
30:00 198.7 F 0.0

Drying Complete

FILLING: 2:51
HEATING: 15:50
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:55:45
CYCLE COMPLETE

07/15/2021 10:48 AM
ACCEPTED: X
REJECTED:
Operator: U02