
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000717
9:26 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:17
STERILIZING 4:24
136.0 C 228kPa 4:24
136.4 C 233kPa 4:54
136.4 C 233kPa 5:24
136.4 C 233kPa 5:54
136.3 C 232kPa 6:24
136.4 C 233kPa 6:54
136.4 C 231kPa 7:24
136.4 C 233kPa 7:54
136.4 C 233kPa 7:54
VENTING 7:54
AIR DRYING 8:49
CYCLE COMPLETE 8:54

Digital Signature #
7D5ABD9D5D662AA8
