
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007298
7:34 02/05/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:01
STERILIZING 3:00
136.5 C 221kPa 3:00
136.3 C 222kPa 3:30
136.4 C 224kPa 4:00
136.2 C 218kPa 4:30
136.3 C 223kPa 5:00
136.3 C 224kPa 5:30
136.4 C 221kPa 6:00
136.3 C 222kPa 6:30
136.3 C 222kPa 6:30
VENTING 6:30
AIR DRYING 7:20
CYCLE COMPLETE 37:20

Digital Signature #
C96D723155CF402F
