

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2637

Name: RITTER 1  
Sterilizer ID: V2178198  
12/05/2022 12:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.7 F 0.0  
2:00 152.3 F 1.0  
4:00 176.2 F 3.6  
6:00 194.6 F 7.3  
8:00 212.9 F 12.1  
10:00 223.5 F 6.8  
12:00 237.4 F 12.0  
14:00 233.9 F 8.8  
16:00 240.7 F 10.5  
18:00 245.4 F 12.9  
20:00 257.0 F 19.2  
22:00 268.2 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.3 F 28.8  
2:00 272.2 F 29.1  
3:00 271.4 F 28.4  
4:00 271.9 F 28.6

Min 271.4 F 28.2  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 232.9 F 0.0  
10:00 228.7 F 0.0  
15:00 226.1 F 0.0  
20:00 231.6 F 0.0  
25:00 227.3 F 0.0  
30:00 224.1 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 23:05  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:03  
CYCLE COMPLETE

12/05/2022 01:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_