
Midmark M11 - v1.0.5
Total Cycles: 4089

Name: RITTER2
Sterilizer ID: V2313228
11/05/2024 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.1 F 0.0
2:00 159.2 F 2.7
4:00 181.2 F 6.7
6:00 202.3 F 11.1
8:00 216.0 F 5.0
10:00 228.2 F 10.0
12:00 232.4 F 8.8
14:00 240.7 F 12.7
16:00 241.2 F 11.7
18:00 253.6 F 18.0
20:00 264.5 F 24.5
22:00 271.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.7 F 30.0
2:00 271.6 F 29.4
3:00 271.1 F 29.5
4:00 272.6 F 30.5

Min 271.0 F 29.2
Max 272.7 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 218.4 F 0.0
10:00 216.2 F 0.0
15:00 220.7 F 0.0
20:00 219.6 F 0.0
25:00 219.6 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

11/05/2024 05:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____