
Midmark M11 - v1.0.0
Total Cycles: 1459

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2021 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.3 F 0.0
2:00 158.3 F 1.6
4:00 169.1 F 4.9
6:00 192.0 F 9.1
8:00 220.8 F 8.8
10:00 221.5 F 7.9
12:00 236.0 F 12.8
14:00 238.1 F 9.8
16:00 235.3 F 8.5
18:00 248.8 F 14.2
20:00 260.5 F 20.7
22:00 271.4 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.6
2:00 272.5 F 29.0
3:00 271.5 F 28.4
4:00 272.2 F 28.8

Min 271.5 F 28.3
Max 273.1 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 239.5 F 0.0
10:00 226.4 F 0.0
15:00 228.9 F 0.0
20:00 227.1 F 0.0
25:00 226.7 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:57
CYCLE COMPLETE

12/15/2021 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____