

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002133  
14:18 07/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:15  
STERILIZING 4:19  
136.0 C 222kPa 4:19  
136.1 C 222kPa 4:49  
136.4 C 225kPa 5:19  
136.2 C 223kPa 5:49  
136.2 C 224kPa 6:19  
136.4 C 222kPa 6:49  
136.3 C 224kPa 7:19  
136.3 C 224kPa 7:49  
136.3 C 224kPa 7:49  
VENTING 7:49  
AIR DRYING 8:42  
CYCLE COMPLETE 11:25

-----  
Digital Signature #  
D881CBE063EF5F40  
-----