
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000186
13:41 06/07/2020
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 5:55
STERILIZING 6:58
136.0 C 220kPa 6:58
136.2 C 222kPa 7:28
136.4 C 222kPa 7:58
136.4 C 225kPa 8:28
136.4 C 225kPa 8:58
136.4 C 223kPa 9:28
136.2 C 219kPa 9:58
136.2 C 223kPa 10:28
136.4 C 221kPa 10:58
136.4 C 224kPa 11:28
136.2 C 222kPa 11:58
136.3 C 222kPa 12:28
136.4 C 224kPa 12:58
136.2 C 222kPa 13:28
136.4 C 224kPa 13:58
136.4 C 224kPa 14:28
136.1 C 223kPa 14:58
136.3 C 222kPa 15:28
136.4 C 223kPa 15:58
136.4 C 223kPa 16:28
CYCLE ABORTED 16:50
STOP BUTTON PRESSED
NOT STERILE
END TIME 16:50

Digital Signature #

BD4982E411B10E84
