
Midmark M11 - v1.0.5
Total Cycles: 1930

Name: RITTER2
Sterilizer ID: V2313228
01/12/2023 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 184.4 F 4.0
4:00 195.5 F 9.2
6:00 215.0 F 7.0
8:00 227.0 F 9.8
10:00 232.7 F 8.5
12:00 243.6 F 13.5
14:00 243.9 F 12.8
16:00 256.8 F 19.6
18:00 268.2 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.0 F 29.8
2:00 272.6 F 30.0
3:00 272.5 F 29.7
4:00 272.1 F 29.5

Min 271.1 F 28.9
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 229.4 F 0.0
10:00 237.5 F 0.0
15:00 228.9 F 0.0
20:00 230.8 F 0.0
25:00 228.2 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

01/12/2023 03:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____