

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001890  
10:31 25/07/2024  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 2:12  
PRESSURIZING 4:14  
STERILIZING 5:02  
136.0 C 222kPa 5:02  
136.2 C 224kPa 5:32  
136.4 C 226kPa 6:02  
136.2 C 224kPa 6:32  
136.3 C 225kPa 7:02  
136.3 C 225kPa 7:32  
136.1 C 222kPa 8:02  
136.4 C 226kPa 8:32  
136.3 C 225kPa 9:02  
136.4 C 226kPa 9:32  
136.4 C 226kPa 10:02  
136.2 C 222kPa 10:32  
136.4 C 227kPa 11:02  
136.2 C 225kPa 11:32  
136.4 C 226kPa 12:02  
136.4 C 225kPa 12:32  
136.4 C 227kPa 13:02  
136.3 C 225kPa 13:32  
136.4 C 226kPa 14:02  
136.1 C 224kPa 14:32  
136.4 C 226kPa 15:02  
136.4 C 226kPa 15:02  
VENTING 15:02  
AIR DRYING 15:53  
CYCLE COMPLETE 18:57  
-----

Digital Signature #  
3970BA7391DD6E2C  
-----