
Midmark M11 - v1.0.0
Total Cycles: 5433

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2025 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 70.1 F 0.4
4:00 84.4 F 1.8
6:00 135.8 F 4.2
8:00 160.1 F 6.9
10:00 171.9 F 10.2
12:00 184.6 F 7.2
14:00 205.0 F 4.5
16:00 216.2 F 7.8
18:00 227.9 F 11.6
20:00 227.0 F 6.6
22:00 239.5 F 11.7
24:00 236.5 F 9.1
26:00 249.7 F 15.4
28:00 260.8 F 21.7
30:00 271.2 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.1
2:00 272.0 F 28.9
3:00 272.5 F 28.9
4:00 272.2 F 29.1

Min 271.6 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 197.5 F 0.0
10:00 204.3 F 0.0
15:00 213.5 F 0.0
20:00 210.1 F 0.0
25:00 213.7 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:04
CYCLE COMPLETE

03/24/2025 09:37 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____