
Midmark M11 - v1.0.0
Total Cycles: 4615

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2024 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 73.9 F 0.3
4:00 113.4 F 0.6
6:00 123.0 F 1.3
8:00 145.5 F 4.1
10:00 173.6 F 7.6
12:00 191.0 F 11.3
14:00 207.2 F 3.3
16:00 217.2 F 6.4
18:00 227.5 F 10.4
20:00 231.6 F 8.3
22:00 239.1 F 11.3
24:00 235.7 F 8.9
26:00 249.8 F 15.4
28:00 261.2 F 21.8
30:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.1
2:00 271.6 F 28.7
3:00 272.4 F 29.1
4:00 271.5 F 28.6

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.4 F 0.0
5:00 212.7 F 0.0
10:00 217.7 F 0.0
15:00 226.5 F 0.0
20:00 225.0 F 0.0
25:00 222.7 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 30:36
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:47
CYCLE COMPLETE

06/07/2024 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____