
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001455
16:04 09/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:26
PRESSURIZING 7:26
STERILIZING 9:28
136.4 C 221kPa 9:28
136.5 C 222kPa 9:58
136.3 C 221kPa 10:28
136.4 C 221kPa 10:59
136.3 C 222kPa 11:29
136.2 C 222kPa 11:59
136.3 C 220kPa 12:29
136.4 C 223kPa 12:58
VENTING 12:59
AIR DRYING 14:06
CYCLE COMPLETE 26:19

Digital Signature #
B9A4C75BD01B066C
