
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000735
7:51 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:02
STERILIZING 3:07
136.4 C 229kPa 3:07
136.3 C 231kPa 3:37
136.3 C 232kPa 4:07
136.1 C 231kPa 4:37
136.3 C 230kPa 5:07
136.2 C 231kPa 5:37
136.3 C 232kPa 6:07
136.2 C 230kPa 6:37
136.2 C 230kPa 6:37
VENTING 6:37
AIR DRYING 7:28
CYCLE COMPLETE 7:35

Digital Signature #
830FACB60F2314EF
