
Midmark M11 - v1.0.5
Total Cycles: 2220

Name: RITTER2
Sterilizer ID: V2313228
04/10/2023 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 138.6 F 1.8
4:00 163.9 F 4.2
6:00 186.4 F 8.2
8:00 207.4 F 13.1
10:00 219.1 F 6.8
12:00 232.9 F 11.9
14:00 232.8 F 8.6
16:00 240.3 F 10.9
18:00 244.4 F 13.0
20:00 256.6 F 19.4
22:00 267.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.1 F 29.6
2:00 272.1 F 29.3
3:00 271.2 F 29.1
4:00 272.7 F 29.8

Min 271.2 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 225.5 F 0.0
10:00 223.5 F 0.0
15:00 221.9 F 0.0
20:00 221.3 F 0.0
25:00 218.8 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

04/10/2023 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____