

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2571

Name: RITTER2  
Sterilizer ID: V2313228  
06/27/2023 07:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.2 F 0.0  
2:00 75.1 F 0.0  
4:00 77.6 F 0.0  
6:00 112.7 F 0.0  
8:00 146.5 F 0.0  
10:00 176.9 F 2.5  
12:00 195.3 F 5.8  
14:00 209.5 F 9.3  
16:00 221.3 F 13.0  
18:00 224.6 F 5.9  
20:00 231.6 F 7.7  
22:00 242.1 F 12.2  
24:00 238.7 F 10.2  
26:00 236.7 F 8.8  
28:00 249.5 F 15.2  
30:00 260.8 F 21.7  
32:00 271.2 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.7 F 29.5  
2:00 271.7 F 28.6  
3:00 272.1 F 29.3  
4:00 272.4 F 29.3  
  
Min 271.4 F 28.5  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.1 F 0.0  
5:00 211.6 F 0.0  
10:00 190.3 F 0.0  
15:00 188.0 F 0.0  
20:00 184.3 F 0.0  
25:00 187.1 F 0.0  
30:00 188.6 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 32:44  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:10:43  
CYCLE COMPLETE

06/27/2023 08:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_