
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000078
12:46 12/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:54
PRESSURIZING 3:55
STERILIZING 5:14
136.4 C 224kPa 5:14
136.2 C 221kPa 5:44
136.3 C 223kPa 6:14
136.4 C 223kPa 6:44
136.4 C 224kPa 7:14
136.4 C 224kPa 7:44
136.4 C 224kPa 8:14
136.4 C 224kPa 8:44
136.4 C 224kPa 8:44
VENTING 8:44
AIR DRYING 9:42
CYCLE COMPLETE 9:45

Digital Signature #
295E72F21292C2E7
