
Midmark M11 - v1.0.5
Total Cycles: 2895

Name: RITTER2
Sterilizer ID: V2313228
09/18/2023 08:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.7 F 0.0
2:00 75.5 F 0.0
4:00 90.2 F 0.0
6:00 113.3 F 0.0
8:00 163.3 F 0.0
10:00 180.6 F 1.4
12:00 210.6 F 4.8
14:00 216.6 F 8.3
16:00 222.9 F 12.0
18:00 226.6 F 6.9
20:00 238.5 F 11.5
22:00 234.1 F 8.5
24:00 242.5 F 11.5
26:00 245.2 F 13.1
28:00 257.0 F 19.4
30:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.1
2:00 272.0 F 29.3
3:00 271.9 F 28.8
4:00 272.4 F 29.6

Min 271.3 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 234.5 F 0.0
10:00 201.6 F 0.0
15:00 222.4 F 0.0
20:00 220.3 F 0.0
25:00 209.1 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

09/18/2023 09:26 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____