

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1012

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2022 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.3 F 0.0  
2:00 196.2 F 2.0  
4:00 210.4 F 7.9  
6:00 227.4 F 12.3  
8:00 234.8 F 10.5  
10:00 234.8 F 8.8  
12:00 239.4 F 10.0  
14:00 249.6 F 15.2  
16:00 261.8 F 22.2  
18:00 272.8 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.7  
1:00 272.5 F 29.5  
2:00 271.7 F 28.6  
3:00 272.6 F 29.5  
4:00 272.1 F 28.8  
  
Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 218.1 F 0.0  
10:00 227.3 F 0.0  
15:00 228.3 F 0.0  
20:00 228.7 F 0.0  
25:00 228.3 F 0.0  
30:00 229.2 F 0.0  
  
Drying Complete

FILLING: 2:22  
HEATING: 18:12  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:44  
CYCLE COMPLETE

04/20/2022 11:58 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_